



# WASHINGTON STATE PATROL SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: April 16, 2006 INITIALS: DEC OWNER: WSP  
 DISTRICT: 8 ADDRESS: \_\_\_\_\_  
 LOCATION: 7 M Shop COUNTY: \_\_\_\_\_  
 MAKE: Haenni TYPE: port wheel weigher ACCURACY CLASS: III  
 MODEL: 101 ID. NO: 24150 CAPACITY: 20,000 lbs SECTIONS: 1  
 MINIMUM GRADUATIONS: 50 lbs REG. ELEMENT: dial PLATFORM: 15" X 26"

| LOAD POSITION | TEST WEIGHT  | SCALE INDICATION | ERROR      | LOAD POSITION | TEST WEIGHT | SCALE INDICATION | ERROR |
|---------------|--------------|------------------|------------|---------------|-------------|------------------|-------|
| <u>Center</u> | <u>5000</u>  | <u>5000</u>      | <u>0</u>   |               |             | <u>-57</u>       |       |
| <u>"</u>      | <u>10000</u> | <u>10050</u>     | <u>+50</u> |               |             | <u>-30</u>       |       |
| <u>"</u>      | <u>15000</u> | <u>15050</u>     | <u>+50</u> |               |             | <u>-47</u>       |       |
| <u>Center</u> | <u>20000</u> | <u>20000</u>     | <u>0</u>   |               |             | <u>-41</u>       |       |
|               |              |                  |            |               |             | <u>-24</u>       |       |
|               |              |                  |            |               |             | <u>00</u>        |       |
|               |              |                  |            |               |             |                  |       |
|               |              |                  |            |               |             |                  |       |
|               |              |                  |            |               |             |                  |       |

NOTE: READINGS OBTAINED FROM AVERAGE OF 2 RUNS

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as is shown on the attached "Certificate of Accuracy", No. 19577

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cramer Date: 4/18/2006

State of Washington

County of Thurston

Signed and sworn to (or affirmed) before me on 4.18.06 by

Michael J. Evans  
(Signature of notary public)

Michael J. EVANS  
Printed Name of Notary

My appointment expires 12-27-07



3000-362-021 (R) (10/98)

DISTRIBUTION: WHITE - COURT GREEN - DISTRICT CANARY - PROPERTY MANAGEMENT PINK & GOLDENROD - SERVICE TRUCKS

Service Technician: Mike Johnson

Form 1 03/31/02